

Infrared Radiometer Calibration Center

of the WRC/PMOD

-IRC-

Julian Gröbner

Tasks within the Workplan

1h) Develop further the establishment of a World Standard Group of absolute longwave radiometers

task: Collaborate with PMOD/WRC on the establishment of a World Standard Group of absolute long-wave radiometers

Deliverable: A World Standard Group of absolute long-wave radiometers established

Deadline May 2006: Not yet achieved!

Tasks within the Workplan

1i) Coordinate the dissemination of the World Radiometric Reference factors to regional and national radiation standards

task: Develop proposal for the dissemination of IR factors to regional and national radiation standards –*centers(?)*

Deliverable: Report to Members

Deadline May 2006: Calibration Center is operational and in use

Overview


- History of the IRC
- Instrumentation (overview)
- Stability of the longwave infrared reference
- Pyrheliometer calibration procedure
- Absolute longwave reference
 - ASR
 - Uncertainty budget
- Current developments / outlook

History

- Official starting date 1 January 2004
 - One Absolute sky scanning Radiometer (ASR)
 - Comprises 4 pyrgeometers
 - 2 modified Eppley PIR, 31463, 31464
 - 2 Kipp&Zonen CG4, FT004, 010535
- „World Infrared Standard Group of pyrgeometers“
- WISG -

This standard group is based on the Absolute Sky-Scanner Radiometer (ASR) developed by R. Philipona

Instrumentation of IRC



Black Body



WISG



ASR

World Infrared Standard Group of pyrgeometers

The WISG coefficients are based on IPASRC-I results with a small adjustment for conditions at PMOD/WRC (Philipona, 2005 private communication).
 September 1999, Great Plains, Oklahoma, US

Table 1. Pyrgeometer Inventory and Calibration and Correction Factors


Serial	Owner ^a	FACT		User		CMDL		PMOD				FIELD					
		T_D	C	C	K	C	K	C	k_1	k_2	k_3	K	C	k_1	k_2	k_3	K
PIR13678F3	BoM Australia	1	4.13	4.29	4.0	3.97	4.0					4.27		0.997			4.0
PIR28492F3	SUNYA USA	1	3.74	3.74	4.0	3.49	4.0					3.43		0.994			4.0
PIR28898F3	University of Maryland USA	1	4.00	4.00	4.0	3.91	3.0					3.90		1.003			3.0
PIR30555F3	NOVA/CMDL USA	1	3.57	3.57	4.0	3.52	3.5					3.69		1.000			3.5
PIR31195F3	NREL USA	1	3.56	3.56	4.0	3.49	3.5					3.41		1.003			3.5
PIR32227F3	Eppley USA	1	3.90	3.90	4.0	3.93	3.5					3.83		1.003			3.5
PIR32690F3	Saudi Arabia	1	4.00	4.00	4.0	4.19	3.5					4.10		1.002			3.5
PIR26036mod	NASA USA	3	(4.14)	3.84	3.48	3.78	4.0	4.36	0.1218	1.0038	3.58	4.42	0.1218	1.0000	3.58		
PIR28146mod	AES Canada	3	(3.88)	3.69	3.42	3.67	4.0	4.34	0.1510	1.0024	3.61	4.32	0.1510	1.0024	3.61		
PIR32305mod	JMA Japan	3	(4.04)	3.92	2.66	3.80	3.5	4.25	0.1112	1.0031	2.75	4.26	0.1112	1.0031	2.75		

^a BoM, Bureau of Meteorology; SUNYA, State University of New York at Albany; SRRB, Surface Radiation Branch; NREL, National Renewable Energy Laboratory; AES, Atmospheric Environmental Service; DOE/ARM, Department of Energy Atmospheric Radiation Measurement Program; DWD, Deutsche Wetter Dienst; PMOD, Physikalisch-Meteorologisches Observatorium Davos; JMA, Japanese Meteorological Agency; K&Z, Kipp & Zonen.


Philipona et al., Atmospheric longwave irradiance uncertainty: Pyrgeometers compared to an absolute sky-scanning radiometer, atmospheric emitted radiance interferometer, and radiative transfer model calculations, JGR, 106, 2001..

Stability

WISG



WISG traceability to IPASRC-I




Stability

Conclusion

- The WISG together with additional instruments have remained stable as a group since the IPASRC-I campaign in September 1999.
- No apparent
 - Drift
 - offset

Pyrgeometer Calibration procedure

- Based on Philipona et al. formula:

$$E = \frac{U_{emf}}{C} (1 + k_1 \sigma T_b^3) + k_2 \sigma T_b^4 - k_3 (T_d^4 - T_b^4)$$



- C, k_1, k_2, k_3 are unknown coefficients and are retrieved from the calibration.

Calibration procedure First Step


- Characterisation in the black body to retrieve k_1, k_2, k_3 . C is retrieved but not used as calibration constant.

Pyregeometer temperature range: +20°C to -10°C

Black Body temperature range: +15°C to -30°C



Results from the regression -> C, k_1, k_2, k_3



Black-Body coefficients C, k1,k2,k3

Typical uncertainties from black-body measurements and retrieval procedure:

$$\delta C/C = \pm 0.02 \quad \rightarrow \quad \delta E/E = \pm 0.01$$


$$\delta k_1 = \pm 0.02$$

$$\delta k_2 = \pm 0.0002 \quad \delta E = \pm 0.1 \text{ W/m}^2 \text{ (T=300 K)}$$

$$\delta k_3 = \pm 0.05 \quad \delta E = \pm 0.3 \text{ W/m}^2 \text{ (T=300 K, } \delta T=1\text{K)}$$

Calibration procedure Second Step

- Comparison to the WISG using atmospheric infrared radiation to determine the pyrgeometer sensitivity C.




Longwave radiation

RMS 0.2 W/m²

+1
W

Sensitivity studies



Calibration procedure Conclusion

- k_1, k_2, k_3 are retrieved from the black body characterisation
- Sensitivity C is obtained from direct comparison to the WISG and thus traceable to ASR.

Table 1. Institutes traceable to the IRC; Calibrations performed since January 2004)

- Deutscher Wetterdienst, Germany
- Eidgenössische Technische Hochschule Zürich, Switzerland
- Météo France, France
- Optimare, Germany
- Alfred-Wegener Institut, Germany
- National Ocean and Air Administration (NOAA), U.S.
- NASA Langley Research Center, U.S.
- Arsenal Research, Austria
- Météo Swiss, Switzerland
- Royal Netherlands Meteorological Institute (KNMI), the Netherlands
- National Renewable Energy Laboratory, U.S.
- Met Office, U. K.
- ENEA, Italy
- Bureau of Meteorology, Australian Government, Australia
- Kipp&Zonen, the Netherlands
- Swedish Meteorological and Hydrological Institute (SMHI), Sweden
- China Meteorological Administration, China

Total 17


Calibration uncertainty based on:

$$u = \text{std}(\text{PIR}) + \text{std}(\text{REF}) + u_{\text{ASR}}$$


Absolute ASR uncertainty estimated at $\pm 1 \text{ W/m}^2$
 (Philipona et al. 2001)

Example:

- $C (*1e-6) = 4.293$
- Irradiance Variation from 175.1 to 354.1 W/m^2
- Residuals from -2.0 to 2.5 W/m^2
- Body Temperature (only night) from -12.26 to 21.39
- Uncertainty REF 0.835 W/m^2 , rel. = 0.33 %
- Uncertainty PIR 0.693 W/m^2 , rel. = 0.27 %
- Relative standard Uncertainty of C = 0.89 %
- Expanded uncertainty (k=2) of C = 0.076 microV/W

Assuming Thermopile = 0.5 mV  $\delta E = 2.1 \text{ W/m}^2$



Ratio between ASR-based and Black-Body based calibration





Spectral mismatch, effect of water in the atmospheric column

Example atmospheric LIR spectra, spectral dome transmissions

Difference test Pyrgometer to WISG



Irradiance measurements – Angular response of pyrgeometers



Absolute sky scanning radiometer (ASR)

Absolute sky scanning radiometer (ASR)


Features:

- Pyroelectric detector
- Flat spectral response
- Calibration based on low temperature black-body
- Irradiance obtained from gaussian integration of 5° sky radiance measurements

Absolute sky scanning radiometer (ASR)

Table 2. Uncertainties due to Errors on ASR Absolute Long-Wave Irradiance Measurements

Uncertainties	Error	Measured Unit (W m^{-2})
Type		
Uncertainty of temperature of blackbody (at 260 K)	$\pm 0.15 \text{ K}$	± 0.6
Uncertainty of chopper temperature during calibration	$\pm 0.05^\circ \text{ K}$	± 0.2
Uncertainty of blackbody emittance	0.999 ± 0.0005	± 0.14
Mismatch due to pyroelectric detector sensitivity	$(100 \text{ W m}^{-2}) \pm 0.5\%$	± 0.5
Field-of-view beam stray light	$(100 \text{ W m}^{-2}) \pm 0.5\%$	± 0.5
Ambient air passage	$(100 \text{ W m}^{-2}) \pm 0.5\%$	± 0.2
Combined standard uncertainty (rss)		± 0.98
Maximum uncertainty (worst case)		± 2.15

$\pm 1 \text{ W m}^{-2}$


Uncertainty budget for ASR as published in Philipona, Appl. Opt., 40, 2001.

Missing: Calibration reproducibility

Comparison blackbody ASR with IRC

nov 05

BB_{ASR}




BB_{IRC}



Comparison blackbody ASR with IRC results

ASR Sensitivity and offset in mV/W


date	ASR	IRC
24 nov	9.19 +121	10.24+31
25 nov	9.57+86	10.26+52
29 nov	9.77-10	
30 nov		10.25-6.5
1 dec		9.84+0
5 dec	9.32+38 9.69-2	9.89+1



Inconsistent results, some unknown systematic?


ASR geometry

Dimensions not to scale!!




„straylight“ from Baffles and walls?

PMOD/WRC Horizon on IRC-Plattform




Sky-Scanning Radiometer (ASR)

Measurements from March, 11, 2005



Sky- or Horizon Temperature
 -10° to -15° C (250-270 Watt)

With $\alpha = 5.4\%$, this gives an uncertainty in the longwave irradiance of 1-3 Watt



Horizon correction at IRC for this specific case: **+2 W \pm 1 W**

ASR / Conclusions

- Concept is sound:
 - Solves problems related to angular and spectral response.
- Prototype instrument
- Manual night-time operation
- Calibration reproducibility seems difficult
- Needs further characterisation

IRC achievements

- Creation of a stable group of pyrgeometers (WISG)
- Development of laboratory characterisations for pyrgeometers to determine:
 - Calibration coefficients (k_1, k_2, k_3) using a well characterised black-body
- Designed calibration procedure to transfer the calibration to test pyrgeometers

ToDo / in Progress

- Construction of a second Black-body (in progress)
- Design and realisation of a second sky scanning radiometer based on an electric substitution pyroelectric detector (in progress)
- Perform comparisons between ASR and WISG
- Angular response measurement (in progress)
- Planning for a spectral response measurement facility for pyrgeometers